

# Notice of Allowability

Application No.

10/550,304

Examiner

Anastasia Midkiff

Applicant(s)

BECKHAUS ET AL.

Art Unit

2882

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address--

All claims being allowable, PROSECUTION ON THE MERITS IS (OR REMAINS) CLOSED in this application. If not included herewith (or previously mailed), a Notice of Allowance (PTOL-85) or other appropriate communication will be mailed in due course. **THIS NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIGHTS.** This application is subject to withdrawal from issue at the initiative of the Office or upon petition by the applicant. See 37 CFR 1.313 and MPEP 1308.

1. ☒ This communication is responsive to Applicant Amendment filed 23 July 2007.
2. ☒ The allowed claim(s) is/are 35, 39 and 40.
3. ☒ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
  - a) ☒ All b) ☐ Some\* c) ☐ None of the:
    1. ☒ Certified copies of the priority documents have been received.
    2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
    3. ☐ Copies of the certified copies of the priority documents have been received in this national stage application from the International Bureau (PCT Rule 17.2(a)).

\* Certified copies not received: \_\_\_\_\_.

Applicant has THREE MONTHS FROM THE "MAILING DATE" of this communication to file a reply complying with the requirements noted below. Failure to timely comply will result in ABANDONMENT of this application.

**THIS THREE-MONTH PERIOD IS NOT EXTENDABLE.**

4. ☐ A SUBSTITUTE OATH OR DECLARATION must be submitted. Note the attached EXAMINER'S AMENDMENT or NOTICE OF INFORMAL PATENT APPLICATION (PTO-152) which gives reason(s) why the oath or declaration is deficient.
  5. ☐ CORRECTED DRAWINGS (as "replacement sheets") must be submitted.
    - (a) ☐ including changes required by the Notice of Draftsperson's Patent Drawing Review (PTO-948) attached
      - 1) ☐ hereto or 2) ☐ to Paper No./Mail Date \_\_\_\_\_.
    - (b) ☐ including changes required by the attached Examiner's Amendment / Comment or in the Office action of Paper No./Mail Date \_\_\_\_\_.
- Identifying indicia such as the application number (see 37 CFR 1.84(c)) should be written on the drawings in the front (not the back) of each sheet. Replacement sheet(s) should be labeled as such in the header according to 37 CFR 1.121(d).
6. ☐ DEPOSIT OF and/or INFORMATION about the deposit of BIOLOGICAL MATERIAL must be submitted. Note the attached Examiner's comment regarding REQUIREMENT FOR THE DEPOSIT OF BIOLOGICAL MATERIAL.

## Attachment(s)

1. ☒ Notice of References Cited (PTO-892)
2. ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
3. ☐ Information Disclosure Statements (PTO/SB/08),  
Paper No./Mail Date \_\_\_\_\_
4. ☐ Examiner's Comment Regarding Requirement for Deposit  
of Biological Material
5. ☐ Notice of Informal Patent Application
6. ☐ Interview Summary (PTO-413),  
Paper No./Mail Date \_\_\_\_\_
7. ☐ Examiner's Amendment/Comment
8. ☒ Examiner's Statement of Reasons for Allowance
9. ☐ Other \_\_\_\_\_

ASM

## **ATTACHMENT TO NOTICE OF ALLOWABILITY**

### ***Allowable Subject Matter***

Claims 35, 39, and 40 are allowed.

With respect to Claims 35 and 39, the prior art of record teaches many of the elements of the claimed invention, including an x-ray system having an image detector built into an x-ray sensitive camera, comprising: an x-ray emitter with a primary diaphragm; a second image detector being provided inside said camera; wherein adjustment means are provided for moving, as desired, said first or second image detectors into proper alignment with an x-ray emitter for the creation of a respective x-ray image; that said adjustment means are provided on casing of said camera; that the adjustment range of said camera is equal to at least one width of said first detector; wherein there is additionally provided an installation for the creation of teleradiographic images; positioning of said camera with respect to said additional installation; and a radiolucent region in an x-ray camera with multiple detectors.

However, prior art fails to teach or fairly suggest the system wherein there is additionally provided an additional detector for the installation of teleradiographic images, so that positioning of said camera includes moving camera out of the path of the additional detector, in the manner required by Claims 35 and 39.

Claim 40 is allowed by virtue of its dependency upon Claim 39.

***Conclusion***

The prior art made of record and not relied upon is considered pertinent to applicant's disclosure.

U.S. Patent to Nyholm (US 7,092,483 B2) regarding an installation with systems for creating both panoramic and teleradiographic (cephalometric) images with x-rays, and a detector with a plurality of sensors for each of said systems, but without a second image detector, positioning said detector with a plurality of sensors out of the path of a second image detector, or a radiolucent region.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Anastasia Midkiff whose telephone number is 571-272-5053. The examiner can normally be reached on M-F 7-4.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Edward Glick can be reached on 571-272-2490. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a



Application/Control Number: 10/550,304

Page 4

Art Unit: 2882

USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

ASM  
8/7/07



EDWARD J. GLICK  
SUPERVISORY PATENT EXAMINER